

COMPREHENSIVE TEST REPORT

Customer Name	G&R COLD FORGING		SERVICE TICKET/ WORK ORDER #	SV1101150935			
Unit Location	SHIPPING B3		Customer #	N/A			
Mfg/Model #	TOLEDO 2256		Serial #	1148075-1AJ			
Capacity	5000	Unit of Measure	16	Number of Divisions	5000	Division Size	16

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP0023IR, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST


Position	As Found	As Left
1	501	500
2	501	500
3	501	500
4	501	500

Weights Applied: 500 16
 Max. Permissible Error: ± 2(d)

<input checked="" type="checkbox"/>	In tolerance with out adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 16	0	0	0	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 16	501	+ 1	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Max. Load*	1000 16	1002	+ 2	± 2	1000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 16	501	+ 1	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Zero	0 16	0	0	± 1	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:			
WEIGHT IDENTIFICATION NUMBERS:		W1 - W30	
Weight Traceability Certificate #:	1359368	CUSTOMER CALIBRATION DATE:	MARCH 2, 2011
		CUSTOMER CALIBRATION DUE:	SEP 5, 2011
PERFORMED BY:	KARL MIKES		
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature

COMPREHENSIVE TEST REPORT

Customer Name	G&R COLD FORGING		SERVICE TICKET/ WORK ORDER #	SV110150935	
Unit Location	SHIPPING B3		Customer #	N/A	
Mfg/Model #	TOLEDO BBA 442		Serial #	2P70613-7LH	
Capacity	70	Unit of Measure	16	Number of Divisions	14000
				Division Size	0.005 16

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP00231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST


Position	As Found	As Left
1	30.005	30.000
2	30.005	30.000
3	30.005	30.000
4	30.005	30.000

Weights Applied: 30 16
 Max. Permissible Error: ± 4(d)

<input checked="" type="checkbox"/>	In tolerance with out adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 16	0.000	0	0	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.000	0	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 16	30.005	+1	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.010	+2	± 6	50.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Max. Load*	70 16	70.015	+3	± 8	70.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.010	+2	± 6	50.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 16	30.005	+1	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.000	0	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 16	0.000	0	± 1	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:			
WEIGHT IDENTIFICATION NUMBERS:		WIP, KIT # 87914	
Weight Traceability Certificate #:	1359368, 1346956	CUSTOMER CALIBRATION DATE:	MARCH 2, 2011
		CUSTOMER CALIBRATION DUE:	SEPT / 2011
PERFORMED BY:	KARL MIKES		
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature

COMPREHENSIVE TEST REPORT

Customer Name	G & R COLD FORGING		SERVICE TICKET/ WORK ORDER #	SV110150925			
Unit Location	Q.C. LAB G3		Customer #	N/A			
Mfg/Model #	METTLER PB1501-S		Serial #	1128033891			
Capacity	1500	Unit of Measure	g	Number of Divisions	15000	Division Size	0.1g

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP00231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST

Position	As Found	As Left
1	500.0	500.0
2	500.0	500.0
3	500.0	500.0
4	500.0	500.0

Weights Applied: 500g
 Max. Permissible Error: ± 3(d)

<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 g	0.0	0	0	0.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 g	50.0	0	± 1	50.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	500 g	500.0	0	± 2	500.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	1000 g	1000.0	0	± 3	1000.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Max. Load*	1500 g	1500.0	0	± 4	1500.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	1000 g	1000.0	0	± 3	1000.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	500 g	500.0	0	± 2	500.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 g	50.0	0	± 1	50.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 g	0.0	0	± 1	0.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:			
WEIGHT IDENTIFICATION NUMBERS:		KIT # 40073	
Weight Traceability Certificate #:	W019564-12998	CUSTOMER CALIBRATION DATE:	MARCH 27 2011
		CUSTOMER CALIBRATION DUE:	SEPT 2011
PERFORMED BY:	KARL MILES		
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature