

COMPREHENSIVE TEST REPORT

Customer Name	G & R Cold Forging		SERVICE TICKET/ WORK ORDER #	SV1107160381			
Unit Location	SHIPPING B1		Customer #	N/A			
Mfg/Model #	HT NIPER SD30		Serial #	2574369.7HE			
Capacity	75	Unit of Measure	16	Number of Divisions	15000	Division Size	

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP0023IR, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST


Position	As Found	As Left
1	30.000	30.000
2	30.000	30.000
3	30.000	30.000
4	30.000	30.000

Weights Applied: 30 16
 Max. Permissible Error: ± 4(d)

<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 16	0.000	0	0	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.000	0	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 16	30.000	0	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.005	+1	± 6	50.000	Y <input type="checkbox"/> N <input type="checkbox"/>
Max. Load*	75 16	75.015	+3	± 8	75.000	Y <input type="checkbox"/> N <input type="checkbox"/>
	50 16	50.005	+1	± 6	50.000	Y <input type="checkbox"/> N <input type="checkbox"/>
	30 16	30.000	0	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.000	0	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 16	0.000	0	± 1	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

- * Maximum load used in test
- No decreasing load applicable

REMARKS:			
WEIGHT IDENTIFICATION NUMBERS:		W22, KIT # 87914	
Weight Traceability Certificate #:	1359368, 1359340	CUSTOMER CALIBRATION DATE:	SEP 23, 2011
PERFORMED BY:		CUSTOMER CALIBRATION DUE: MARCH 2012	
KARL MILES			
Technician Name (Please print)		Technician Signature	
Where Applicable:		Customer Signature	
Customer Name (Please print)			

COMPREHENSIVE TEST REPORT

Customer Name	GFR COLD FORGING			SERVICE TICKET/ WORK ORDER #	SV11071603P1		
Unit Location	SHIPPING G1			Customer #	W1919		
Mfg/Model #	MT 2256			Serial #	1118906-7BE		
Capacity	5000	Unit of Measure	16	Number of Divisions	5000	Division Size	16

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VPO0231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST

Position	As Found	As Left
1	500	500
2	500	500
3	500	500
4	500	500

Weights Applied: 500 16
 Max. Permissible Error: ±2(d)

<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 16	0	0	0	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 16	500	0	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Max. Load*	1000 16	1000	0	± 2	1000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 16	500	0	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Zero	0 16	0	0	± 1	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:

WEIGHT IDENTIFICATION NUMBERS:	W1 - W30		
Weight Traceability Certificate #:	1359368	CUSTOMER CALIBRATION DATE:	SEP. 23, 2011
		CUSTOMER CALIBRATION DUE:	MAR/2012
PERFORMED BY:	KARL MIKER		Karl M. G.
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature



COMPREHENSIVE TEST REPORT

Customer Name	GER Cold Forging		SERVICE TICKET/ WORK ORDER #	SV11071C03R1			
Unit Location	IN PLANT G1		Customer #	N/A			
Mfg/Model #	DIBIWEIGH DWP 121		Serial #	82464A 07070085			
Capacity	400	Unit of Measure	lb	Number of Divisions	20 000	Division Size	0.02 16

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VPO0231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST

Position	As Found	As Left
1	20.00	20.00
2	20.00	20.00
3	20.00	20.00
4	20.00	20.00

Weights Applied: 20 16
 Max. Permissible Error: ± 2 (d)

<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 16	0.00	0	0	0.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.00	0	± 1	10.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	20 16	20.00	0	± 2	20.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.02	+1 0 R.H.	± 3	50.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Max. Load*	100 16	100.04	+2	± 4	100.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 16	50.02	+1	± 3	50.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	20 16	20.00	0	± 2	20.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 16	10.00	0	± 1	10.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 16	0.00	0	± 1	0.00	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:

WEIGHT IDENTIFICATION NUMBERS:	W11 W14 KIT # 87914		
Weight Traceability Certificate #:	1359368, 1359340	CUSTOMER CALIBRATION DATE:	SEP 23 2011
		CUSTOMER CALIBRATION DUE:	MARCH 2012
PERFORMED BY:	KARL MIKES		
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature