

COMPREHENSIVE TEST REPORT

Customer Name	G&R Cold Forging		SERVICE TICKET/ WORK ORDER #	SV1107160381			
Unit Location	SHIPPING G3		Customer #	N/A			
Mfg/Model #	MT 2256		Serial #	1148075-1AJ			
Capacity	5000	Unit of Measure	lb	Number of Divisions	5000	Division Size	1 lb

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VP00231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST

Position	As Found	As Left
1	500	500
2	500	500
3	500	500
4	500	500

Weights Applied: 500 lb
 Max. Permissible Error: ± 2(d)

<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 lb	0	0	0	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 lb	500	0	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Max. Load*	1000 lb	1000	0	± 2	1000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
	500 lb	500	0	± 1	500	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
						Y <input type="checkbox"/> N <input type="checkbox"/>
Zero	0 lb	0	0	± 1	0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:

WEIGHT IDENTIFICATION NUMBERS: W1 - W30

Weight Traceability Certificate #: 1359368 CUSTOMER CALIBRATION DATE: SEP. 23, 2011 CUSTOMER CALIBRATION DUE: MAR/2012

PERFORMED BY: KARL MIKES [Signature]
 Technician Name (Please print) Technician Signature

Where Applicable: Customer Name (Please print) Customer Signature

COMPREHENSIVE TEST REPORT

Customer Name	C & R Cold Forging			SERVICE TICKET/ WORK ORDER #	SV1107160381		
Unit Location	SHIPPING G3			Customer #	N/A		
Mfg/Model #	MT BBA 442			Serial #	2870613-7LH		
Capacity	70	Unit of Measure	lb	Number of Divisions	14000	Division Size	0.00516

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VPO0231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST


Position	As Found	As Left
1	30.010	30.000
2	30.010	30.000
3	30.010	30.000
4	30.010	30.000

Weights Applied: 30 lb
 Max. Permissible Error: ± 4(d)

<input checked="" type="checkbox"/>	In tolerance with adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 lb	0.000	0	0	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 lb	10.005	+1	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 lb	30.010	+2	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 lb	50.020	+4	± 6	50.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Max. Load*	70 lb	70.025	+5	± 8	70.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 lb	50.020	+4	± 6	50.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	30 lb	30.010	+2	± 4	30.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	10 lb	10.005	+1	± 2	10.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 lb	0.000	0	± 1	0.000	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test
 No decreasing load applicable

REMARKS:			
WEIGHT IDENTIFICATION NUMBERS:		W14, KIT # 87914	
Weight Traceability Certificate #:	1359362, 1359340	CUSTOMER CALIBRATION DATE:	SEPT 23, 2011
		CUSTOMER CALIBRATION DUE:	MAR / 2012
PERFORMED BY:	KARL MIKES		
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature

COMPREHENSIVE TEST REPORT

Customer Name	G & R Cold Forging		SERVICE TICKET/ WORK ORDER #	SV1107160381	
Unit Location	Q.C. LAB G3		Customer #	N/A	
Mfg/Model #	MT PB 1501 S		Serial #	1128033891	
Capacity	1500	Unit of Measure	g	Number of Divisions	15000
				Division Size	0.19

The weighing equipment noted on this report has been tested and/or calibrated in compliance with METTLER TOLEDO procedure VPO0231R, and NIST Handbook 44 and/or Canada Standard where applicable.

SHIFT TEST

Position	As Found	As Left
1	500.1	500.0
2	500.1	500.0
3	500.1	500.0
4	500.1	500.0

Weights Applied:

500g

Max. Permissible Error:

±3(d)


<input checked="" type="checkbox"/>	In tolerance without adjustment
<input type="checkbox"/>	In tolerance after adjustment
<input type="checkbox"/>	Out of tolerance
<input type="checkbox"/>	Shift test not applicable

Test Load	Weights Applied	As Found	Error: Plus or Minus (d)	Allowable Error (d)	As Left	Within Tolerance? Y/N
Zero	0 g	0.0	0	0	0.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 g	50.0	0	± 1	50.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	500 g	500.1	+ 1	± 2	500.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	1000 g	1000.2	+ 2	± 3	1000.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Max. Load*	1500 g	1500.3	+ 3	± 4	1500.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	1000 g	1000.2	+ 2	± 3	1000.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	500 g	500.1	+ 1	± 2	500.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
	50 g	50.0	0	± 1	50.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>
Zero	0 g	0.0	0	± 1	0.0	Y <input checked="" type="checkbox"/> N <input type="checkbox"/>

* Maximum load used in test

No decreasing load applicable

REMARKS:

WEIGHT IDENTIFICATION NUMBERS:	KIT # 42672		
Weight Traceability Certificate #:	135962P	CUSTOMER CALIBRATION DATE:	SEP 23, 2011
		CUSTOMER CALIBRATION DUE:	MAR/2012
PERFORMED BY:	KARL MILLES		
	Technician Name (Please print)		Technician Signature
Where Applicable:	Customer Name (Please print)		Customer Signature