



**ABSOLUTE MEASUREMENT SYSTEMS**  
**CMM Calibration & Service**

3986 Bentridge Road  
 Mississauga, ON L5N 7V7  
 Phone: (905) 785-7600

Certificate Number: C-0208

Reissue Number: \_\_\_\_\_  
 If applicable



# Calibration Certificate

Company: Windsor Machine & Stamping (2009) Ltd.  
 Address: 7084 Smith Industrial Drive  
 City: McGregor  
 Province: Ontario  
 Postal Code: N0R 1J0

CMM: CMM Software  
 Serial No.:  
 Gauge No.:  
 Technician:

IMS Impact II  
 Virtual DMS  
 600072  
 Robert Schaff

Procedure: B89.4.1b-2001  
 Units: Millimeters  
 Temp. Range: 20.8 °C - Max 20.2 °C - Min  
 Date Calibrated: March 12 2013  
 Date Issued: May 8 2013

Repeatability		
Specification	As Found	After Correction
X	0.001	N/A
Y	0.001	N/A
Z	0.001	N/A
		Uncertainty 1.2 µm 1.2 µm 1.2 µm

Squareness**		
Specification	As Found	After Correction
XY	-0.0003 degrees	N/A
YZ	0 degrees	N/A
ZX	-0.0002 degrees	N/A
		Uncertainty (0.8 + 1.5) µm (0.8 + 1.5) µm (0.8 + 1.5) µm

Linear Accuracy		
Specification	As Found	After Correction
X	0.001	N/A
Y	0.001	N/A
Z	0.001	N/A
		Uncertainty (0.3 + 2.5) µm (0.3 + 2.5) µm (0.3 + 2.5) µm

Volumetric Performance		
Specification	As Found	After Correction
N/A	0.010	N/A
		Uncertainty (0.8 + 1.5) µm

This Calibration Certificate certifies that the instrument listed above has been calibrated in accordance with applicable national, international, and AMS specifications and standards. This Certificate of Calibration is issued in accordance with the applicable requirements of ISO/IEC 17025:2005. Calibration results are traceable to the National Research Council (NRC), the National Institute of Standards and Technology (NIST), or the National Physical Laboratory (NPL), where applicable. The above noted CMM was calibrated at the Customer's location. The calibration results relate only to the CMM noted above, and for the environmental and instrument conditions at the time of calibration. This Calibration Certificate is part of a multi-page document containing 8 pages.

**Traceability Information:**

Length Reference Standard: Optodyne Laser (MCV 500) - S.N. 0807003080, S.N. 0807002607  
 Environmental Thermometer: Fluke Digital Thermometer - S.N. 97690209  
 Ballbar Spheres: Renishaw 25.000 mm - S.N. 139481, S.N. 139485  
 Ball bar length: 366.8 mm

\*After Correction results are only reported for coordinate measuring machines whose volumetric parameters have been adjusted.  
 \*\* Squareness measurements are not part of AMS scope of accreditation, but are taken for reference purposes (if requested).

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
 \*This certificate may only be reproduced in its entirety.

**Measurement Uncertainty:**  
 The stated expanded uncertainties are given at a coverage factor of k=2 for a level of confidence of approximately 95% assuming a normal distribution.

Name: Robert Schaff  
 Title: Service Manager

Signature: \_\_\_\_\_

Form: 1.0/SOCALCERT Rev. 1.10  
 Date: January 29, 2012  
 Prepared by: Sarah Sullivan Approved by: Thomas Pegios



Repeatability 'As Found'

Meas. #	X Axis	Y Axis	Z Axis
1	0.0018	-0.0009	-0.0008
2	0.0014	-0.0010	-0.0008
3	0.0017	-0.0009	-0.0005
4	0.0018	-0.0008	-0.0008
5	0.0016	-0.0009	-0.0009
6	0.0012	-0.0011	-0.0006
7	0.0016	-0.0006	-0.0010
8	0.0016	-0.0008	-0.0011
9	0.0016	-0.0008	-0.0008
10	0.0018	-0.0005	-0.0009

Max: 0.0018 -0.0005 -0.0005  
 Min: 0.0012 -0.0011 -0.0011  
 Repeatability: 0.0006 0.0006 0.0006

Repeatability 'After Correction'

Meas. #	X Axis	Y Axis	Z Axis
1	N/A	N/A	N/A
2	N/A	N/A	N/A
3	N/A	N/A	N/A
4	N/A	N/A	N/A
5	N/A	N/A	N/A
6	N/A	N/A	N/A
7	N/A	N/A	N/A
8	N/A	N/A	N/A
9	N/A	N/A	N/A
10	N/A	N/A	N/A

Max: N/A N/A N/A  
 Min: N/A N/A N/A  
 Repeatability: N/A N/A N/A

Units: Millimeters

Volumetric Performance 'As Found'

Position	Trial 1	Trial 2	Trial 3	Average
1	366.801	N/A	N/A	366.801
2	366.802	N/A	N/A	366.802
3	366.799	N/A	N/A	366.799
4	366.799	N/A	N/A	366.799
5	366.794	N/A	N/A	366.794
6	366.796	N/A	N/A	366.796
7	366.800	N/A	N/A	366.800
8	366.801	N/A	N/A	366.801
9	366.802	N/A	N/A	366.802
10	366.792	N/A	N/A	366.792
11	366.793	N/A	N/A	366.793
12	366.801	N/A	N/A	366.801
13	N/A	N/A	N/A	N/A
14	N/A	N/A	N/A	N/A
15	N/A	N/A	N/A	N/A
16	N/A	N/A	N/A	N/A
17	N/A	N/A	N/A	N/A
18	N/A	N/A	N/A	N/A
19	N/A	N/A	N/A	N/A
20	N/A	N/A	N/A	N/A
21	N/A	N/A	N/A	N/A
22	N/A	N/A	N/A	N/A
23	N/A	N/A	N/A	N/A
24	N/A	N/A	N/A	N/A
25	N/A	N/A	N/A	N/A
26	N/A	N/A	N/A	N/A
27	N/A	N/A	N/A	N/A
28	N/A	N/A	N/A	N/A
29	N/A	N/A	N/A	N/A
30	N/A	N/A	N/A	N/A
31	N/A	N/A	N/A	N/A
32	N/A	N/A	N/A	N/A
33	N/A	N/A	N/A	N/A
34	N/A	N/A	N/A	N/A
35	N/A	N/A	N/A	N/A

Max: 366.802  
 Min: 366.792  
 Volumetric: 0.010

Volumetric Performance 'After Correction'

Position	Trial 1	Trial 2	Trial 3	Average
1	N/A	N/A	N/A	N/A
2	N/A	N/A	N/A	N/A
3	N/A	N/A	N/A	N/A
4	N/A	N/A	N/A	N/A
5	N/A	N/A	N/A	N/A
6	N/A	N/A	N/A	N/A
7	N/A	N/A	N/A	N/A
8	N/A	N/A	N/A	N/A
9	N/A	N/A	N/A	N/A
10	N/A	N/A	N/A	N/A
11	N/A	N/A	N/A	N/A
12	N/A	N/A	N/A	N/A
13	N/A	N/A	N/A	N/A
14	N/A	N/A	N/A	N/A
15	N/A	N/A	N/A	N/A
16	N/A	N/A	N/A	N/A
17	N/A	N/A	N/A	N/A
18	N/A	N/A	N/A	N/A
19	N/A	N/A	N/A	N/A
20	N/A	N/A	N/A	N/A
21	N/A	N/A	N/A	N/A
22	N/A	N/A	N/A	N/A
23	N/A	N/A	N/A	N/A
24	N/A	N/A	N/A	N/A
25	N/A	N/A	N/A	N/A
26	N/A	N/A	N/A	N/A
27	N/A	N/A	N/A	N/A
28	N/A	N/A	N/A	N/A
29	N/A	N/A	N/A	N/A
30	N/A	N/A	N/A	N/A
31	N/A	N/A	N/A	N/A
32	N/A	N/A	N/A	N/A
33	N/A	N/A	N/A	N/A
34	N/A	N/A	N/A	N/A
35	N/A	N/A	N/A	N/A

Max: N/A  
 Min: N/A  
 Volumetric: N/A

Squareness\*\* 'As Found'

	XY		YZ		ZX	
	+X +Y	+X -Y	+Y +Z	+Y -Z	+Z +X	+Z -X
Trial 1	366.801	366.799	366.799	366.799	366.799	366.798
Trial 2	N/A	N/A	N/A	N/A	N/A	N/A
Trial 3	N/A	N/A	N/A	N/A	N/A	N/A
Average:	366.801	366.799	366.799	366.799	366.799	366.798
Angle:	-0.0003 degrees		0 degrees		-0.0002 degrees	
Squareness:	1.57080 arc sec.		1.57080 arc sec.		1.57080 arc sec.	
Range:	0.002		0.000		0.001	

	XY		YZ		ZX	
	+X +Y	+X -Y	+Y +Z	+Y -Z	+Z +X	+Z -X
Trial 1	N/A	N/A	N/A	N/A	N/A	N/A
Trial 2	N/A	N/A	N/A	N/A	N/A	N/A
Trial 3	N/A	N/A	N/A	N/A	N/A	N/A
Average:	N/A	N/A	N/A	N/A	N/A	N/A
Angle:	N/A		N/A		N/A	
Squareness:	N/A		N/A		N/A	
Range:	N/A		N/A		N/A	

Squareness\*\* 'After Correction'

Reissue Number: \_\_\_\_\_

if applicable

Form: 1.0ISOCALCERT Rev. 1.10

Prepared by: Sarah Sullivan Approved by: Thomas Feiglos

Date: January 29, 2012



**ABSOLUTE MEASUREMENT SYSTEMS**  
CMM Calibration & Service

Certificate Number: C-0208

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Date Calibrated: March 12 2013

Reissue Number: \_\_\_\_\_

if applicable

X Axis

Cell	Location	Trial 1	Trial 2	Trial 3	Average
1	0	0.000	0.000	0.000	0.000
2	75	0.001	0.001	0.001	0.001
3	150	0.001	0.000	0.001	0.001
4	225	0.001	0.001	0.001	0.001
5	300	0.000	0.001	0.000	0.000
6	375	0.001	0.001	0.001	0.001
7	450	0.001	0.001	0.001	0.001
8	525	0.001	0.001	0.001	0.001
9	600	0.000	0.000	0.000	0.000
10	675	0.000	0.000	0.000	0.000
11	750	0.001	0.001	0.001	0.001

Linear Displacement - As Found

Range					
Axis	Trial 1	Trial 2	Trial 3	Average	
X Axis	Max 0.001	Max 0.000	Max 0.001	Max 0.001	Min 0.000
	Min 0.000	Min 0.001	Min 0.000	Min 0.001	Max 0.000

Linear Displacement Summary	
Axis	As Found
X Axis	Result 0.001

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
\*This certificate may only be reproduced in its entirety.



**ABSOLUTE MEASUREMENT SYSTEMS**  
 CMM Calibration & Service

Certificate Number: C-0208

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Reissue Number: \_\_\_\_\_  
 if applicable

Date Calibrated: March 12 2013

Cell	Location	Trial 1	Trial 2	Trial 3	Average
1	0	0.000	0.000	0.000	0.000
2	75	0.000	0.001	0.000	0.000
3	150	0.000	0.000	0.000	0.000
4	225	0.000	0.000	0.000	0.000
5	300	0.000	0.001	0.000	0.000
6	375	0.001	0.001	0.001	0.001
7	450	0.000	0.001	0.000	0.000
8	525	0.000	0.000	0.000	0.000
9	600	0.000	0.000	0.000	0.000
10	675	0.000	0.000	0.000	0.000
11	750	0.001	0.001	0.001	0.001

Linear Displacement - As Found

Range					
Axis	Trial 1	Trial 2	Trial 3	Average	
Y Axis	Max 0.001 Min 0.000	Max 0.001 Min 0.000	Max 0.001 Min 0.000	Max 0.001 Min 0.000	Max 0.001 Min 0.000

Linear Displacement Summary	
Axis	Result
Y Axis	0.001

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
 \*This certificate may only be reproduced in its entirety.





**ABSOLUTE MEASUREMENT SYSTEMS**  
CMM Calibration & Service

Certificate Number: C-0208

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Date Calibrated: March 12 2013

Reissue Number: \_\_\_\_\_

if applicable

Z Axis

Cell	Location	Trial 1	Trial 2	Trial 3	Average
1	0	0.000	0.000	0.000	0.000
2	55	0.000	0.000	0.000	0.000
3	110	0.000	0.000	0.000	0.000
4	165	0.000	0.000	0.000	0.000
5	220	0.000	0.000	0.000	0.000
6	275	0.001	0.000	0.000	0.000
7	330	-0.001	-0.001	-0.001	-0.001
8	385	0.001	0.000	0.000	0.000
9	440	-0.001	-0.001	-0.001	-0.001
10	495	0.001	0.000	0.000	0.000

Linear Displacement - As Found

Range					
Axis	Trial 1	Trial 2	Trial 3	Average	
Z Axis	Max 0.001	Max 0.000	Max 0.000	Max 0.000	Max 0.000
	Min -0.001	Min -0.001	Min -0.001	Min -0.001	Min -0.001

Linear Displacement Summary	
Axis	As Found
Z Axis	Result
	0.001

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
\*This certificate may only be reproduced in its entirety.



**ABSOLUTE MEASUREMENT SYSTEMS**  
CMM Calibration & Service

Certificate Number: C-0208

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Date Calibrated: March 12 2013

Reissue Number: \_\_\_\_\_

if applicable

X Axis Cell

Location	Trial 1	Trial 2	Trial 3	Average
1	0			
2	75			
3	150			
4	225			
5	300			
6	375			
7	450			
8	525			
9	600			
10	675			
11	750			

Linear Displacement - After Correction

Range					
Axis	Trial 1	Trial 2	Trial 3	Average	
Max	N/A	N/A	N/A	N/A	N/A
Min	N/A	N/A	N/A	N/A	N/A
Max	N/A	N/A	N/A	N/A	N/A
Min	N/A	N/A	N/A	N/A	N/A

Linear Displacement Summary	
After Correction	
Axis	X Axis
Result	N/A

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
\*This certificate may only be reproduced in its entirety.



**ABSOLUTE MEASUREMENT SYSTEMS**  
CMM Calibration & Service

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Date Calibrated: March 12 2013

Reissue Number: \_\_\_\_\_  
if applicable

Certificate Number: C-0208


**Linear Displacement - After Correction**

Cell	Y Axis	Location	Trial 1	Trial 2	Trial 3	Average
1	0					
2	75					
3	150					
4	225					
5	300					
6	375					
7	450					
8	525					
9	600					
10	675					
11	750					

Range						
Axis	Trial 1	Trial 2	Trial 3	Average	Max	Min
Y Axis	N/A	N/A	N/A	N/A	N/A	N/A

Linear Displacement Summary	
Axis	Y Axis
After Correction	Result
	N/A

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
\*This certificate may only be reproduced in its entirety.



**ABSOLUTE MEASUREMENT SYSTEMS**  
CMM Calibration & Service

Certificate Number: C-0208

Company: Windsor Machine & Stamping (2009) Ltd.

Serial Number: 600072

Date Calibrated: March 12 2013

Reissue Number: \_\_\_\_\_  
if applicable

Z Axis  
Cell

Location	Trial 1	Trial 2	Trial 3	Average
10	495			
9	440			
8	385			
7	330			
6	275			
5	220			
4	165			
3	110			
2	55			
1	0			

Linear Displacement - After Correction

Range				
Z Axis	Max	Min	Max	Min
Axis	Trial 1	Trial 2	Trial 3	Average
	N/A	N/A	N/A	N/A
	Max	Max	Max	Max
	Min	Min	Min	Min

Linear Displacement Summary	
Axis	After Correction
Z Axis	Result
	N/A

\*The results relate only to this CMM, environmental and instrument conditions at the time of calibration.  
\*This certificate may only be reproduced in its entirety.